

### **Prilikom merenja temperature termoparom nema rezultata**

Temperaturna sonda tipa K u izvedbi Z3431-X na izlazu očitava 1 mV po °C.

For measurements this means:

Connect the probe to the voltage measurement sockets and select the mV DC measuring range. The following results, for example, are obtained after measurement: 24.5 mV which corresponds to 24.5° C.

The number is correct. The unit of measure just has to be changed mentally. This applies to the STARLINE range as well!